

KISUMU WATER AND SEWERAGE COMPANY LTD



01.08.2018

ADDENDA NO. 001

This is an addenda for tender No. KWSC/12/2018-2019

The following are further specifications as per item numbers:

- **ITEM 166 Hot Plate Magnetic Stirrer**
 - Top plate dimensions (140mm/5.5" diameter)
 - Digital stirrer preferred but analogue can also pass.

- **ITEM 169 TU5 Turbidity Meter**

Operating within a range of 0-50°C and not necessarily come with an RFID

- **ITEM 174 Oven - 270 degrees for Microbial Analysis**

Standard laboratory size

- **ITEM 186 Distiller**

20 litres per hour distiller

- **ITEM 167 Autoclave**

Vessel volume= 0.024m³
Rated working pressure= 0.14-0.16MPa

- **ITEM 182 Microscope (BD-SW50 Infinity Semi apochromatic Biological Microscope)**

Optical magnification= 40X-1000X
Light source = 12V/1000W halogen lamp
View tube = 30° Tilt, 360°Rotating
Eye piece = High eye level WF PL10 × (Φ25mm)

Other terms and conditions of the tender remain unchanged.